

To
The Director PME
OUAT
Bhubaneswar

Sub: Corrigendum to Tender NO. 361 / CAET, Dt. 22nd Feb, 2020

Sir,

Due to typographical error three items for which valid financial bids have been received are again floated in the above tender notice. I request you to kindly put a corrigendum deleting the following items from the above tender notice.

10. Hand hold Ultrasonic homogeniser

Suitable for manual operation or for stand use Power: 50 watts Frequency: 30kHz Automatic frequency tuning system Amplitude adjustable 20-100% Pulse 0-100% Dry running protection Timer (0-100 min) Stand, mounting tools and sound protection box With Sonotrode made of titanium suitable for samples from 10ml up to 250ml

11. Ultrasonic bath


Capacity of the bath: 3 litres Ultrasonic power~W: 480 Ultrasonic nominal power: 120W Max. frequency: 35kHz Min. heating/cooling temperature: 30°C Max. heating/cooling temperature: 80°C Heating power~W: 140 Protection class: IP32

12. Temperature and humidity cabinet Cabinet

interior size: 150 liters (approx) (SS 304) Double door (interior glass and outer SS) Temperature range: 10 to 60 deg C with ± 1 deg Capacitive type humidity sensor Digitally controlled active humidification and dehumidification 40% to 95% RH Microprocessor based PID digital temp indicator cum controller Condensation free

Thank you and with regards

Yours faithfully


Dr. Manoj Kumar Panda
Chairman, Gr-A (Laboratory Equipments)