



ODISHA UNIVERSITY OF AGRICULTURE & TECHNOLOGY
BHUBANESWAR-3

SHORT QUOTATION CALL NOTICE OF EQUIPMENT

Required Specification of shaking machine indented by AICRP Micronutrient, OUAT

Designed for shaking Adjustable volume of conical flasks (100 ml-250 ml) of different size by **adjusting** compartments. Fitted with heavy duty motor to work on 220V, AC 50 Hz single phase. The rate of motion is adjustable mechanically by belt tension. Speed is fixed in same models to approximately 275. Size of Carrier: 36"x18"x 4" Strokes per Minutes: 150 to 250 Motor: **2 HP**
Body Type: covered, Uninterrupted shaking for 12 hours compulsory.

Terms and Conditions:

1. The supplier/Agency firm must provide the printed literature indicating photographs and detailed technical specification, manual and their functions.
2. It is necessary that specific brand, make, model etc. shall be clearly mentioned in the quotation.
3. A letter indicating that-
 - a) The quoted item is new and has not been tampered.
 - b) The manufacturer shall provide three years on site guarantee/warranty, after sales, service during warranty/ comprehensive annual maintainance contract/ annual maintainance contract period. If the authorized dealer/bidder fails to provide the same.
4. The price should be quoted in Indian currency including on site delivery price to different offices/ departments of OUAT, Bhubaneswar
5. Attested certificate of PAN and GST IN showing the registered number, latest GST return and photograph of the supplier should be provided.
6. The quotation should be send **To Prof. S.K. Pattanayak, Chairman, Local Purchase Committee College of Agriculture, OUAT, Bhubaneswar-751003.** With in 5 days of publication of the notice by speed post/courier post.

Sd/-Dr. S.K. Pattanayak
Chairman, Local Purchase Committee,
C.A., OUAT, Bhubaneswar