



Department of Fruit Science & Horticulture Technology
College of Agriculture
Orissa University of Agriculture and Technology
Bhubaneswar – 751 003

From : Prof. D.K. Dash
Prof. and Head, Dept. of FSHT

No. 31 / FSHT
Date 15.01.18

QUOTATION CALL NOTICE

Sealed quotations are invited from interested firms having valid PAN/GST number for supply of following equipments to this office.

Sl. No.	Equipment	Specifications
1	Thermohygrometer	Humidity : Range – 25-95%, Resolution – 1%, Accuracy - $\pm 2\%$ at 60%RH Temperature : Range – 32-122°F, Resolution – 1°F, Accuracy - $\pm 2^\circ\text{F}$ Response time : 30 seconds to 3.5 minutes. Warranty – 3years
2	Lux meter	Range – 0 to 4,00,000 lux, Resolution – 0.01lux, Accuracy - $\pm 3\%$ rdg $\pm 0.5\%$ full scale Warranty – 3years

These equipments are to be delivered at the office of the undersigned. The quotations should be superscribed as “**quotation for supply of hygrometer and lux meter**” and reach the office of the undersigned on or before 24.1.2018 (5PM).

(Head, Dept. of FSHT)

Copy to Notice board (College of Agril, University) for wider circulation.
Copy to Director, PME for display of this notice in OUAT Website.