



BIOTECHNOLOGY CUM TISSUE CULTURE CENTRE

ORISSA UNIVERSITY OF AGRICULTURE & TECHNOLOGY
BHUBANESWAR-751 003, ODISHA

Fax No:-
0674 2397780
Phone no:-
0674 2354720
Email: -
btcc.2009.ouat@gmail.com

Dr(Mrs) Sashikala Beura

Director

No. 824 /TC

Dated: 9/2/21

QUOTATION CALL NOTICE


Sealed quotations are invited from interested firms/vendors having valid PAN/GST number for supply and installation of following equipments/item in the Biotechnology-cum-Tissue Culture Centre, OUAT, Bhubaneswar. The details of items with specifications are given below.


Sl. No.	Equipment/Item	Specifications
1.	Glass bead sterilizer	Temperature 230° C-250° C Power 220-240 volts A.C. Crucible size: 2" Dia and 6" Depth Compact-suitable for any Laminar Air Flow Bench
2.	pH meter, combined electrode	Range (pH) : 0 to 14 Resolution (pH) : 0.1/0.01 Accuracy (pH) : ±0.01 pH Temperature compensation : Automatic or manual Display : LCD Power : 220-240 V AC, 50/60 Hz
3.	Microwave oven	LG make, 15L capacity
4.	Electronic balance	Maximum Capacity : 100 gm. Readability : 0.1 mg, Repeatability : 0.1 mg. AC Adapter 230 Volts. Tare range full, Internal Calibration

The quotations, along with PAN/GST clearance certificate, in sealed cover, should reach the Director, Biotechnology-cum-Tissue Culture Centre, OUAT, Bhubaneswar-751003 through speed post/courier only. The quotations should be superscribed as "Quotation for supply and installation of small equipments in BTCC" and reach the office of the undersigned on or before 16.02.2021 (5PM).

Memo No. 825 /Flori. Dt. 9/2/21

Copy submitted to Director, PME for display of this notice in OUAT Website.


(DIRECTOR)


(DIRECTOR)

Copy to Notice board, BTCC, OUAT, Bhubaneswar